

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120318\
 Data File : VN052640.D
 Acq On : 3 Dec 2018 19:10
 Operator : MD\SY
 Sample : PB115231ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115231ZHE#11

Quant Time: Dec 04 01:30:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1199128	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1849589	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1612429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	589363	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	622806	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
35) Dibromofluoromethane	7.59	113	570034	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
50) Toluene-d8	10.09	98	2253052	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	12.40	95	691066	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
Target Compounds						
16) Acetone	3.82	43	15224	4.421	ug/l #	85
20) Methylene Chloride	4.55	84	119448	8.864	ug/l	97
43) Isopropyl Acetate	8.17	43	66366	3.679	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120318\
Data File : VN052640.D
Acq On : 3 Dec 2018 19:10
Operator : MD\SY
Sample : PB115231ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115231ZHE#11

Quant Time: Dec 04 01:30:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

