

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120318\
 Data File : VN052628.D
 Acq On : 3 Dec 2018 13:34
 Operator : MD\SY
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Dec 04 00:39:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	91	0.00
2 M	Dichlorodifluoromethane	1.709	1.638	4.2	85	0.00
3 M	Chloromethane	2.266	2.170	4.2	87	0.00
4 M	Vinyl Chloride	2.287	2.241	2.0	89	0.00
5 M	Bromomethane	1.304	1.353	-3.8	96	0.00
6 M	Chloroethane	1.360	1.352	0.6	89	0.00
7 M	Trichlorofluoromethane	2.823	2.826	-0.1	91	0.00
8 T	Diethyl Ether	1.042	1.079	-3.6	97	0.00
9	1,1,2-Trichlorotrifluoroeth	1.789	1.783	0.3	88	0.00
10 M	1,1-Dichloroethene	1.682	1.696	-0.8	92	0.00
11	Methyl Iodide	2.336	2.529	-8.3	102	0.00
12	Methyl Acetate	1.429	1.606	-12.4	102	0.00
13 M	Acrolein	0.168	0.143	14.9	84	0.00
14 M	Acrylonitrile	0.609	0.653	-7.2	95	0.00
15 M	Acetone	0.158	0.189	-19.6	100	0.00
16 M	Carbon Disulfide	5.326	4.979	6.5	86	0.00
17	Allyl chloride	2.934	2.667	9.1	82	0.00
18 M	Methylene Chloride	1.948	1.988	-2.1	94	0.00
19 M	trans-1,2-Dichloroethene	1.810	1.822	-0.7	93	0.00
20 T	Diisopropyl ether	6.166	6.179	-0.2	91	0.00
21 M	1,1-Dichloroethane	3.514	3.556	-1.2	93	0.00
22 M	cis-1,2-Dichloroethene	2.075	2.120	-2.2	93	0.00
23 M	tert-Butyl Alcohol	0.091	0.111	-22.0	108	0.00
24 M	Methyl tert-Butyl Ether	4.413	4.763	-7.9	98	0.00
25 M	Chloroform	3.390	3.465	-2.2	93	0.00
26	Cyclohexane	3.314	3.177	4.1	87	0.00
27 s	1,2-Dichloroethane-d4	2.044	2.077	-1.6	92	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
29	1,1-Dichloropropene	0.458	0.449	2.0	91	0.00
30 M	2-Butanone	0.123	0.136	-10.6	100	0.00
31	2,2-Dichloropropane	0.422	0.428	-1.4	96	0.00
32 M	1,1,1-Trichloroethane	0.477	0.481	-0.8	95	0.00
33 M	Carbon Tetrachloride	0.417	0.414	0.7	93	0.00
34 M	Benzene	1.377	1.369	0.6	92	0.00
35	Methacrylonitrile	0.152	0.132	13.2	80	0.00
36 M	1,2-Dichloroethane	0.417	0.421	-1.0	94	0.00
37 M	Trichloroethene	0.349	0.347	0.6	93	0.00
38	Methylcyclohexane	0.567	0.539	4.9	87	0.00
39 M	1,2-Dichloropropane	0.369	0.355	3.8	90	0.00
40	Dibromomethane	0.200	0.208	-4.0	96	0.00
41 M	Bromodichloromethane	0.441	0.436	1.1	93	0.00
42 M	Vinyl Acetate	0.663	0.660	0.5	93	0.00
43	Ethyl Acetate	0.261	0.272	-4.2	94	0.00
44	Isopropyl Acetate	0.457	0.553	-21.0	114	0.00
45 T	1,4-Dioxane	0.003	0.003#	0.0	108	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120318\
 Data File : VN052628.D
 Acq On : 3 Dec 2018 13:34
 Operator : MD\SY
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Dec 04 00:39:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.237	0.249	-5.1	99	0.00
47	n-amyl Acetate	0.379	0.387	-2.1	100	0.00
48 M	t-1,3-Dichloropropene	0.427	0.433	-1.4	99	0.00
49 T	cis-1,3-Dichloropropene	0.521	0.514	1.3	94	0.00
50 M	1,1,2-Trichloroethane	0.283	0.296	-4.6	95	0.00
51	Ethyl methacrylate	0.355	0.379	-6.8	103	0.00
52	1,3-Dichloropropane	0.503	0.514	-2.2	94	0.00
53 M	Dibromochloromethane	0.311	0.325	-4.5	97	0.00
54 M	1,2-Dibromoethane	0.276	0.291	-5.4	98	0.00
55 M	2-Chloroethyl vinyl ether	0.208	0.219	-5.3	97	0.00
56 M	Bromoform	0.204	0.217	-6.4	102	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
58 M	4-Methyl-2-Pentanone	0.291	0.318	-9.3	98	0.00
59 M	2-Hexanone	0.199	0.221	-11.1	99	0.00
60 S	4-Bromofluorobenzene	0.459	0.457	0.4	92	0.00
61 M	Tetrachloroethene	0.369	0.392	-6.2	96	0.00
62 M	Toluene	1.654	1.674	-1.2	92	0.00
63 S	Toluene-d8	1.391	1.375	1.2	90	0.00
64 M	Chlorobenzene	1.009	1.010	-0.1	93	0.00
65	1,1,1,2-Tetrachloroethane	0.350	0.357	-2.0	96	0.00
66 M	Ethyl Benzene	1.782	1.774	0.4	92	0.00
67 M	m/p-Xylenes	0.664	0.670	-0.9	92	0.00
68 M	o-Xylene	0.645	0.661	-2.5	95	0.00
69 M	Styrene	1.040	1.045	-0.5	93	0.00
70	Isopropylbenzene	1.715	1.737	-1.3	93	0.00
71 M	1,1,2,2-Tetrachloroethane	0.373	0.404	-8.3	98	0.00
72	1,2,3-Trichloropropane	0.320	0.356	-11.2	108	0.00
73	Bromobenzene	0.415	0.427	-2.9	96	0.00
74	n-propylbenzene	1.998	1.958	2.0	89	0.00
75	2-Chlorotoluene	1.160	1.172	-1.0	92	0.00
76	1,3,5-Trimethylbenzene	1.385	1.410	-1.8	93	0.00
77	t-1,4-Dichloro-2-butene	0.101	0.105	-4.0	101	0.00
78	4-Chlorotoluene	1.158	1.151	0.6	91	0.00
79	tert-butylbenzene	1.206	1.245	-3.2	94	0.00
80	1,2,4-Trimethylbenzene	1.386	1.418	-2.3	92	0.00
81	sec-Butylbenzene	1.680	1.684	-0.2	91	0.00
82	p-Isopropyltoluene	1.431	1.436	-0.3	92	0.00
83 M	1,3-Dichlorobenzene	0.729	0.731	-0.3	93	0.00
84 M	1,4-Dichlorobenzene	0.705	0.714	-1.3	94	0.00
85	n-Butylbenzene	1.229	1.160	5.6	87	0.00
86 T	Hexachloroethane	0.242	0.235	2.9	91	0.00
87 M	1,2-Dichlorobenzene	0.694	0.724	-4.3	95	0.00
88	1,2-Dibromo-3-Chloropropane	0.052	0.058	-11.5	105	0.00
89	1,2,4-Trichlorobenzene	0.342	0.339	0.9	95	0.00
90	Hexachlorobutadiene	0.224	0.228	-1.8	88	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120318\
Data File : VN052628.D
Acq On : 3 Dec 2018 13:34
Operator : MD\SY
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Dec 04 00:39:42 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 27 00:48:33 2018
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	0.691	0.703	-1.7	98	0.00
92	1,2,3-Trichlorobenzene	0.315	0.330	-4.8	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0