

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120320\
 Data File : VN064902.D
 Acq On : 3 Dec 2020 16:14
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 04 04:12:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	179209	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	321197	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	328857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	148391	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	107230	51.47	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	102.94%	
35) Dibromofluoromethane	7.55	113	99577	51.81	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	103.62%	
50) Toluene-d8	10.06	98	395943	50.42	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	100.84%	
62) 4-Bromofluorobenzene	12.38	95	157156	55.92	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	111.84%	

Target Compounds

					Qvalue	
52) Toluene	10.12	92	3184	0.567	ug/l	99
84) 1,2,4-Trimethylbenzene	13.01	105	2669	0.308	ug/l	92
95) Naphthalene	15.11	128	3606	0.401	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120320\
Data File : VN064902.D
Acq On : 3 Dec 2020 16:14
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 04 04:12:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

