

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120320\
 Data File : VN064904.D
 Acq On : 3 Dec 2020 17:02
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 04 04:20:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	179199	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	315409	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	322226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	141153	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	106138	50.95	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.90%	
35) Dibromofluoromethane		7.55	113	96778	51.28	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.56%	
50) Toluene-d8		10.06	98	387454	50.24	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.48%	
62) 4-Bromofluorobenzene		12.38	95	150862	54.66	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	7870	11.492	ug/l	94
20) Methylene Chloride	4.51	84	2904	1.353	ug/l #	93
52) Toluene	10.12	92	4322	0.783	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120320\
Data File : VN064904.D
Acq On : 3 Dec 2020 17:02
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 04 04:20:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

