

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120320\
 Data File : VN064905.D
 Acq On : 3 Dec 2020 17:26
 Operator : JC/MD
 Sample : L4938-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IB1-TREATED-EFFLU

Quant Time: Dec 04 13:36:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	175943	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	311505	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	323582	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	142177	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	106670	52.15	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	104.30%	
35) Dibromofluoromethane	7.55	113	96226	51.62	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	103.24%	
50) Toluene-d8	10.06	98	391498	51.41	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	102.82%	
62) 4-Bromofluorobenzene	12.37	95	152025	55.78	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	111.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120320\
Data File : VN064905.D
Acq On : 3 Dec 2020 17:26
Operator : JC/MD
Sample : L4938-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IB1-TREATED-EFFLU

Quant Time: Dec 04 13:36:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

