

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069753.D
 Acq On : 4 Dec 2021 6:55
 Operator : JC/MD
 Sample : M4922-19MSD
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 04 07:22:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1056609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1775376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1660582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	718209	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	777595	47.664	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.320%
35) Dibromofluoromethane	8.021	113	537974	48.376	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.760%
50) Toluene-d8	10.440	98	2124342	47.575	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.140%
62) 4-Bromofluorobenzene	12.728	95	773884	47.839	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.680%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.073	85	638095	51.778	ug/l	100
3) Chloromethane	2.303	50	717480	46.471	ug/l	99
4) Vinyl Chloride	2.448	62	700217	49.323	ug/l	98
5) Bromomethane	2.850	94	327267	42.432	ug/l	99
6) Chloroethane	3.019	64	421533	48.455	ug/l	96
7) Trichlorofluoromethane	3.379	101	906556	48.655	ug/l	98
8) Diethyl Ether	3.824	74	356754	49.714	ug/l	83
9) 1,1,2-Trichlorotrifluo...	4.207	101	551740	52.826	ug/l	95
10) Methyl Iodide	4.419	142	593037	41.563	ug/l	97
11) Tert butyl alcohol	5.377	59	602427	195.355	ug/l	99
12) 1,1-Dichloroethene	4.183	96	525597	52.222	ug/l	88
13) Acrolein	4.041	56	424499	161.647	ug/l	97
14) Allyl chloride	4.840	41	1092113	48.155	ug/l	95
15) Acrylonitrile	5.554	53	1849952	222.628	ug/l	98
16) Acetone	4.285	43	1744162	177.728	ug/l	97
17) Carbon Disulfide	4.529	76	1354535	52.094	ug/l	100
18) Methyl Acetate	4.854	43	1154183	38.964	ug/l	93
19) Methyl tert-butyl Ether	5.615	73	1933329	49.403	ug/l	97
20) Methylene Chloride	5.095	84	614909	47.134	ug/l #	85
21) trans-1,2-Dichloroethene	5.597	96	543653	49.732	ug/l	86
22) Diisopropyl ether	6.495	45	2243024	50.601	ug/l	93
23) Vinyl Acetate	6.433	43	8079419	224.034	ug/l	94
24) 1,1-Dichloroethane	6.388	63	1165595	50.761	ug/l	98
25) 2-Butanone	7.335	43	2710562	203.284	ug/l	91
26) 2,2-Dichloropropane	7.327	77	731099	40.950	ug/l	97
27) cis-1,2-Dichloroethene	7.324	96	675330	51.132	ug/l	88
28) Bromochloromethane	7.656	49	579821	48.558	ug/l #	79
29) Tetrahydrofuran	7.683	42	1791270	221.236	ug/l	88
30) Chloroform	7.817	83	1164384	50.210	ug/l	100
31) Cyclohexane	8.094	56	1095617	48.258	ug/l	90
32) 1,1,1-Trichloroethane	8.013	97	990292	50.132	ug/l	95
36) 1,1-Dichloropropene	8.220	75	846819	50.105	ug/l	96
37) Ethyl Acetate	7.412	43	993876	39.510	ug/l	95
38) Carbon Tetrachloride	8.206	117	867213	52.325	ug/l	100
39) Methylcyclohexane	9.459	83	992075	49.349	ug/l	90
40) Benzene	8.461	78	2545701	49.278	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069753.D
 Acq On : 4 Dec 2021 6:55
 Operator : JC/MD
 Sample : M4922-19MSD
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 04 07:22:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.635	41	526349	44.574	ug/l	90
42) 1,2-Dichloroethane	8.531	62	1018640	48.846	ug/l	96
43) Isopropyl Acetate	8.560	43	1652294	44.386	ug/l #	95
44) Trichloroethene	9.215	130	2764187	224.181	ug/l	90
45) 1,2-Dichloropropane	9.488	63	689983	49.737	ug/l	99
46) Dibromomethane	9.577	93	448763	49.325	ug/l	92
47) Bromodichloromethane	9.762	83	874747	50.781	ug/l	99
48) Methyl methacrylate	9.558	41	813575	46.832	ug/l	89
49) 1,4-Dioxane	9.569	88	249825	801.466	ug/l #	82
51) 4-Methyl-2-Pentanone	10.330	43	5600920	231.931	ug/l	95
52) Toluene	10.505	92	1581229	50.135	ug/l	98
53) t-1,3-Dichloropropene	10.717	75	914515	44.461	ug/l	99
54) cis-1,3-Dichloropropene	10.188	75	980650	45.290	ug/l #	89
55) 1,1,2-Trichloroethane	10.896	97	626983	49.270	ug/l	99
56) Ethyl methacrylate	10.759	69	1055841	44.248	ug/l #	87
57) 1,3-Dichloropropane	11.044	76	1112195	49.636	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.059	63	627	18.478	ug/l #	45
59) 2-Hexanone	11.084	43	3999819	218.952	ug/l	93
60) Dibromochloromethane	11.237	129	630810	46.442	ug/l	100
61) 1,2-Dibromoethane	11.344	107	645965	49.009	ug/l	99
64) Tetrachloroethene	10.977	164	520787	46.252	ug/l	96
65) Chlorobenzene	11.768	112	1660299	49.284	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.840	131	601531	50.731	ug/l	100
67) Ethyl Benzene	11.843	91	3092278	50.532	ug/l	96
68) m/p-Xylenes	11.953	106	2268287	101.803	ug/l	92
69) o-Xylene	12.280	106	1134813	50.564	ug/l	92
70) Styrene	12.294	104	1838684	51.666	ug/l	95
71) Bromoform	12.457	173	453352	44.992	ug/l #	100
73) Isopropylbenzene	12.578	105	2975257	46.561	ug/l	98
74) N-amyl acetate	12.387	43	1151705	40.755	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.827	83	982176	44.449	ug/l	99
76) 1,2,3-Trichloropropane	12.878	75	1192485	57.897	ug/l	75
77) Bromobenzene	12.859	156	688172	45.850	ug/l	79
78) n-propylbenzene	12.918	91	3405115	47.887	ug/l	96
79) 2-Chlorotoluene	13.007	91	2079987	46.293	ug/l	93
80) 1,3,5-Trimethylbenzene	13.058	105	2433427	47.313	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.623	75	244287	39.726	ug/l	94
82) 4-Chlorotoluene	13.103	91	2000404	47.342	ug/l	93
83) tert-Butylbenzene	13.323	119	2114866	46.438	ug/l	93
84) 1,2,4-Trimethylbenzene	13.366	105	2386736	48.727	ug/l	98
85) sec-Butylbenzene	13.500	105	2912663	47.151	ug/l	97
86) p-Isopropyltoluene	13.613	119	2331317	48.282	ug/l	96
87) 1,3-Dichlorobenzene	13.616	146	1159987	47.043	ug/l	96
88) 1,4-Dichlorobenzene	13.694	146	1125604	46.451	ug/l	97
89) n-Butylbenzene	13.940	91	1843840	47.074	ug/l	98
90) Hexachloroethane	14.208	117	404664	47.734	ug/l	88
91) 1,2-Dichlorobenzene	13.989	146	1133784	47.191	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.603	75	177064	44.345	ug/l	78
93) 1,2,4-Trichlorobenzene	15.262	180	502264	43.049	ug/l	98
94) Hexachlorobutadiene	15.370	225	279413	43.883	ug/l	98
95) Naphthalene	15.504	128	1398885	40.664	ug/l	99
96) 1,2,3-Trichlorobenzene	15.700	180	492299	43.216	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069753.D
Acq On : 4 Dec 2021 6:55
Operator : JC/MD
Sample : M4922-19MSD
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Dec 04 07:22:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.086 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 1056609
Ion Ratio Lower Upper
168 100
99 60.7 41.3 61.9



#2
Dichlorodifluoromethane
Concen: 51.778 ug/l
RT: 2.073 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 85 Resp: 638095
Ion Ratio Lower Upper
85 100
87 32.6 16.3 48.9









#3
Chloromethane
Concen: 46.471 ug/l
RT: 2.303 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 717480
Ion Ratio Lower Upper
50 100
52 32.3 25.4 38.0



#4
Vinyl Chloride
Concen: 49.323 ug/l
RT: 2.448 min Scan# 171
Delta R.T. 0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 62 Resp: 700217
Ion Ratio Lower Upper
62 100
64 31.3 25.8 38.6









#5
Bromomethane
Concen: 42.432 ug/l
RT: 2.850 min Scan# 319
Delta R.T. 0.013 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 327267
Ion Ratio Lower Upper
94 100
96 93.0 73.8 110.8



#6
Chloroethane
Concen: 48.455 ug/l
RT: 3.019 min Scan# 384
Delta R.T. 0.011 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 64 Resp: 421533
Ion Ratio Lower Upper
64 100
66 30.8 26.3 39.5









#7
 Trichlorofluoromethane
 Concen: 48.655 ug/l
 RT: 3.379 min Scan# 515
 Delta R.T. 0.011 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55




Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:101 Resp: 906556
 Ion Ratio Lower Upper
 101 100
 103 64.1 50.1 75.1



#8
 Diethyl Ether
 Concen: 49.714 ug/l
 RT: 3.824 min Scan# 684
 Delta R.T. 0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55

Tgt Ion: 74 Resp: 356754
 Ion Ratio Lower Upper
 74 100
 45 110.6 47.0 141.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.826 ug/l
RT: 4.207 min Scan# 81
Delta R.T. 0.005 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 551740
Ion Ratio Lower Upper
101 100
85 46.2 37.2 55.8
151 75.9 65.9 98.9



#10
Methyl Iodide
Concen: 41.563 ug/l
RT: 4.419 min Scan# 906
Delta R.T. 0.005 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion:142 Resp: 593037
Ion Ratio Lower Upper
142 100
127 46.6 35.4 53.0
141 14.5 11.6 17.4





#11

Tert butyl alcohol

Concen: 195.355 ug/l

RT: 5.377 min Scan# 11

Delta R.T. -0.013 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 59 Resp: 602427

Ion Ratio Lower Upper

59 100

57 10.7 8.7 13.1



#12

1,1-Dichloroethene

Concen: 52.222 ug/l

RT: 4.183 min Scan# 818

Delta R.T. 0.008 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55




Tgt Ion: 96 Resp: 525597

Ion Ratio Lower Upper

96 100

61 186.0 131.0 196.6

98 65.6 52.4 78.6









#13
Acrolein
Concen: 161.647 ug/l
RT: 4.041 min Scan# 765
Delta R.T. 0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 424499
Ion Ratio Lower Upper
56 100
55 72.9 56.6 85.0



#14
Allyl chloride
Concen: 48.155 ug/l
RT: 4.840 min Scan# 1063
Delta R.T. 0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 41 Resp: 1092113
Ion Ratio Lower Upper
41 100
39 73.4 61.2 91.8
76 27.2 26.0 39.0









#15
Acrylonitrile
Concen: 222.628 ug/l
RT: 5.554 min Scan# 11
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 53 Resp: 1849952
Ion Ratio Lower Upper
53 100
52 84.6 66.2 99.2
51 37.1 29.6 44.4



#16
Acetone
Concen: 177.728 ug/l
RT: 4.285 min Scan# 856
Delta R.T. -0.005 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 43 Resp: 1744162
Ion Ratio Lower Upper
43 100
58 27.3 23.1 34.7









#17
Carbon Disulfide
Concen: 52.094 ug/l
RT: 4.529 min Scan# 945
Delta R.T. 0.005 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 1354535
Ion Ratio Lower Upper
76 100
78 9.2 7.5 11.3



#18
Methyl Acetate
Concen: 38.964 ug/l
RT: 4.854 min Scan# 1068
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 43 Resp: 1154183
Ion Ratio Lower Upper
43 100
74 20.6 19.4 29.2





#19

Methyl tert-butyl Ether

Concen: 49.403 ug/l

RT: 5.615 min Scan# 1135

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 73 Resp: 1933329

Ion Ratio Lower Upper

73 100

57 22.5 16.8 25.2



#20

Methylene Chloride

Concen: 47.134 ug/l

RT: 5.095 min Scan# 1158

Delta R.T. 0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Tgt Ion: 84 Resp: 614909




Ion Ratio Lower Upper


84 100

49 141.0 93.5 140.3#

51 43.3 29.4 44.2


86 63.0 53.0 79.4








#21
trans-1,2-Dichloroethene
Concen: 49.732 ug/l
RT: 5.597 min Scan# 11
Delta R.T. 0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 96 Resp: 543653
Ion Ratio Lower Upper
96 100
61 154.8 105.8 158.6
98 63.3 52.0 78.0



#22
Diisopropyl ether
Concen: 50.601 ug/l
RT: 6.495 min Scan# 1680
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 45 Resp: 2243024
Ion Ratio Lower Upper
45 100
43 60.1 51.7 77.5
87 23.8 23.2 34.8
59 10.4 9.4 14.0









#23
 Vinyl Acetate
 Concen: 224.034 ug/l
 RT: 6.433 min Scan# 1
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 8079419
 Ion Ratio Lower Upper
 43 100
 86 8.5 8.5 12.7



#24
 1,1-Dichloroethane
 Concen: 50.761 ug/l
 RT: 6.388 min Scan# 1640
 Delta R.T. 0.000 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55

Tgt Ion: 63 Resp: 1165595
 Ion Ratio Lower Upper
 63 100
 98 6.0 3.3 9.9
 100 3.6 2.1 6.5





#25

2-Butanone

Concen: 203.284 ug/l

RT: 7.335 min Scan# 1993

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 2710562

Ion Ratio Lower Upper

43 100

72 22.2 21.7 32.5



#26

2,2-Dichloropropane

Concen: 40.950 ug/l

RT: 7.327 min Scan# 1990

Delta R.T. 0.000 min

Lab File: VN069753.D




Acq: 4 Dec 2021 6:55


Tgt Ion: 77 Resp: 731099

Ion Ratio Lower Upper

77 100

97 21.9 11.6 34.8









#27
 cis-1,2-Dichloroethene
 Concen: 51.132 ug/l
 RT: 7.324 min Scan# 1989
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 96 Resp: 675330
 Ion Ratio Lower Upper
 96 100
 61 157.4 0.0 275.0
 98 63.6 0.0 129.6



#28
 Bromochloromethane
 Concen: 48.558 ug/l
 RT: 7.656 min Scan# 2113
 Delta R.T. 0.000 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55

Tgt Ion: 49 Resp: 579821
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 4.0
 130 65.0 67.8 101.8#





#29

Tetrahydrofuran

Concen: 221.236 ug/l

RT: 7.683 min Scan# 21

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55





Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 42 Resp: 1791270

Ion	Ratio	Lower	Upper
42	100		
72	37.5	36.3	54.5
71	36.1	34.7	52.1



#30

Chloroform

Concen: 50.210 ug/l

RT: 7.817 min Scan# 2173




Delta R.T. -0.003 min


Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Tgt Ion: 83 Resp: 1164384

Ion	Ratio	Lower	Upper
83	100		
85	64.7	51.9	77.9





#31

Cyclohexane

Concen: 48.258 ug/l

RT: 8.094 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 56 Resp: 1095617

Ion Ratio Lower Upper

56 100

69 32.5 27.1 40.7

84 79.8 73.4 110.0



#32

1,1,1-Trichloroethane

Concen: 50.132 ug/l

RT: 8.013 min Scan# 2246

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55




Tgt Ion: 97 Resp: 990292

Ion Ratio Lower Upper

97 100

99 63.8 50.1 75.1

61 50.0 35.2 52.8





#33

1,2-Dichloroethane-d4

Concen: 47.664 ug/l

RT: 8.437 min Scan# 2405

Delta R.T. 0.000 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 777595

Ion Ratio Lower Upper

65 100

67 51.0 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.965 min Scan# 2601

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55




Tgt Ion: 114 Resp: 1775376


Ion Ratio Lower Upper

114 100

63 23.1 0.0 37.8


88 16.1 0.0 27.4








#35
Dibromofluoromethane
Concen: 48.376 ug/l
RT: 8.021 min Scan# 21
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:113 Resp: 537974
Ion Ratio Lower Upper
113 100
111 102.7 82.2 123.2
192 18.5 16.9 25.3




#36
1,1-Dichloropropene
Concen: 50.105 ug/l
RT: 8.220 min Scan# 2323
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55



Tgt Ion: 75 Resp: 846819
Ion Ratio Lower Upper
75 100
110 32.2 18.4 55.2
77 30.7 24.4 36.6





#37

Ethyl Acetate

Concen: 39.510 ug/l

RT: 7.412 min Scan# 2023

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 43 Resp: 993876

Ion Ratio Lower Upper

43 100

61 12.1 11.5 17.3

70 8.2 7.9 11.9



#38

Carbon Tetrachloride

Concen: 52.325 ug/l

RT: 8.206 min Scan# 2318

Delta R.T. 0.000 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55




Tgt Ion: 117 Resp: 867213


Ion Ratio Lower Upper

117 100

119 95.1 76.1 114.1


121 30.8 24.2 36.2








#39
Methylcyclohexane
Concen: 49.349 ug/l
RT: 9.459 min Scan# 21
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 83 Resp: 992075
Ion Ratio Lower Upper
83 100
55 83.8 58.1 87.1
98 44.1 38.4 57.6



#40
Benzene
Concen: 49.278 ug/l
RT: 8.461 min Scan# 2413
Delta R.T. 0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55


Tgt Ion: 78 Resp: 2545701
Ion Ratio Lower Upper
78 100
77 23.9 18.6 27.8








#41
Methacrylonitrile
Concen: 44.574 ug/l
RT: 7.635 min Scan# 2106
Delta R.T. -0.005 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	41	Resp:	526349
Ion Ratio	Lower	Upper	
41	100		
39	64.7	48.1	72.1
67	72.4	68.3	102.5
52	32.5	29.6	44.4



#42
1,2-Dichloroethane
Concen: 48.846 ug/l
RT: 8.531 min Scan# 2439
Delta R.T. 0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55


Tgt Ion:	62	Resp:	1018640
Ion Ratio	Lower	Upper	
62	100		
98	7.5	0.0	17.6








#43
Isopropyl Acetate
Concen: 44.386 ug/l
RT: 8.560 min Scan# 2451
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 1652294
Ion Ratio Lower Upper
43 100
61 23.4 20.4 30.6
87 10.8 11.2 16.8#



#44
Trichloroethene
Concen: 224.181 ug/l
RT: 9.215 min Scan# 2694
Delta R.T. -0.002 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion:130 Resp: 2764187
Ion Ratio Lower Upper
130 100
95 105.3 0.0 191.2





#45

1,2-Dichloropropane

Concen: 49.737 ug/l

RT: 9.488 min Scan# 21

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 689983

Ion Ratio Lower Upper

63 100

65 30.6 24.7 37.1



#46

Dibromomethane

Concen: 49.325 ug/l

RT: 9.577 min Scan# 2829

Delta R.T. 0.000 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55




Tgt Ion: 93 Resp: 448763


Ion Ratio Lower Upper

93 100

95 81.3 65.9 98.9

174 92.1 84.6 127.0









#47
Bromodichloromethane
Concen: 50.781 ug/l
RT: 9.762 min Scan# 2898
Delta R.T. -0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 874747
Ion Ratio Lower Upper
83 100
85 64.1 50.9 76.3
127 7.9 7.1 10.7



#48
Methyl methacrylate
Concen: 46.832 ug/l
RT: 9.558 min Scan# 2822
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 41 Resp: 813575
Ion Ratio Lower Upper
41 100
69 76.3 72.4 108.6
39 68.4 58.4 87.6





#49

1,4-Dioxane

Concen: 801.466 ug/l

RT: 9.569 min Scan# 21

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 88 Resp: 249825

Ion Ratio Lower Upper

88 100

43 38.1 23.4 35.2#

58 78.1 50.6 75.8#



#50

Toluene-d8

Concen: 47.575 ug/l

RT: 10.440 min Scan# 3151

Delta R.T. -0.003 min

Lab File: VN069753.D




Acq: 4 Dec 2021 6:55


Tgt Ion: 98 Resp: 2124342

Ion Ratio Lower Upper

98 100

100 64.3 50.4 75.6









#51
4-Methyl-2-Pentanone
Concen: 231.931 ug/l
RT: 10.330 min Scan# 3110
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 5600920
Ion Ratio Lower Upper
43 100
58 35.1 30.4 45.6



#52
Toluene
Concen: 50.135 ug/l
RT: 10.505 min Scan# 3175
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 92 Resp: 1581229
Ion Ratio Lower Upper
92 100
91 175.1 142.0 213.0









#53
t-1,3-Dichloropropene
Concen: 44.461 ug/l
RT: 10.717 min Scan# 31
Delta R.T. -0.002 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 914515
Ion Ratio Lower Upper
75 100
77 31.7 25.6 38.4



#54
cis-1,3-Dichloropropene
Concen: 45.290 ug/l
RT: 10.188 min Scan# 3057
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 75 Resp: 980650
Ion Ratio Lower Upper
75 100
77 30.7 26.4 39.6
39 64.4 42.5 63.7#









#55
1,1,2-Trichloroethane
Concen: 49.270 ug/l
RT: 10.896 min Scan# 31
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 97 Resp: 626983
Ion Ratio Lower Upper
97 100
83 87.2 67.8 101.6
85 56.9 45.4 68.2
99 61.7 49.4 74.0



#56
Ethyl methacrylate
Concen: 44.248 ug/l
RT: 10.759 min Scan# 3270
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 69 Resp: 1055841
Ion Ratio Lower Upper
69 100
41 78.0 51.4 77.0#
39 47.7 34.6 52.0





#57

1,3-Dichloropropane

Concen: 49.636 ug/l

RT: 11.044 min Scan# 31

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 1112195

Ion Ratio Lower Upper

76 100

78 31.9 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 18.478 ug/l

RT: 10.059 min Scan# 3009

Delta R.T. 0.016 min

Lab File: VN069753.D




Acq: 4 Dec 2021 6:55


Tgt Ion: 63 Resp: 627

Ion Ratio Lower Upper

63 100

106 0.0 23.9 35.9#









#59
2-Hexanone
Concen: 218.952 ug/l
RT: 11.084 min Scan# 31
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 3999819
Ion Ratio Lower Upper
43 100
58 48.5 26.9 80.7



#60
Dibromochloromethane
Concen: 46.442 ug/l
RT: 11.237 min Scan# 3448
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion:129 Resp: 630810
Ion Ratio Lower Upper
129 100
127 77.1 38.6 116.0









#61
1,2-Dibromoethane
Concen: 49.009 ug/l
RT: 11.344 min Scan# 3488
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 645965
Ion Ratio Lower Upper
107 100
109 93.9 75.7 113.5



#62
4-Bromofluorobenzene
Concen: 47.839 ug/l
RT: 12.728 min Scan# 4004
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 95 Resp: 773884
Ion Ratio Lower Upper
95 100
174 75.8 0.0 187.2
176 72.5 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.744 min Scan# 3637
Delta R.T. -0.002 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:117 Resp: 1660582
Ion Ratio Lower Upper
117 100
82 60.2 42.4 63.6
119 32.4 25.4 38.2



#64
Tetrachloroethene
Concen: 46.252 ug/l
RT: 10.977 min Scan# 3351
Delta R.T. -0.002 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion:164 Resp: 520787
Ion Ratio Lower Upper
164 100
166 128.0 101.9 152.9
129 96.7 72.0 108.0
131 94.7 70.9 106.3









#65
Chlorobenzene
Concen: 49.284 ug/l
RT: 11.768 min Scan# 3646
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 1660299
Ion Ratio Lower Upper
112 100
114 32.0 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 50.731 ug/l
RT: 11.840 min Scan# 3673
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion:131 Resp: 601531
Ion Ratio Lower Upper
131 100
133 95.3 47.9 143.7
119 64.6 32.1 96.5





#67

Ethyl Benzene

Concen: 50.532 ug/l

RT: 11.843 min Scan# 3

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 91 Resp: 3092278

Ion Ratio Lower Upper

91 100

106 29.7 25.7 38.5



#68

m/p-Xylenes

Concen: 101.803 ug/l

RT: 11.953 min Scan# 3715

Delta R.T. 0.000 min

Lab File: VN069753.D




Acq: 4 Dec 2021 6:55


Tgt Ion:106 Resp: 2268287

Ion Ratio Lower Upper

106 100

91 213.4 160.6 240.8









#69
o-Xylene
Concen: 50.564 ug/l
RT: 12.280 min Scan# 3837
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 1134813
Ion Ratio Lower Upper
106 100
91 224.8 105.8 317.3



#70
Styrene
Concen: 51.666 ug/l
RT: 12.294 min Scan# 3842
Delta R.T. -0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion:104 Resp: 1838684
Ion Ratio Lower Upper
104 100
78 54.9 38.7 58.1
103 56.4 45.0 67.6









#71
Bromoform
Concen: 44.992 ug/l
RT: 12.457 min Scan# 3903
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 453352
Ion Ratio Lower Upper
173 100
175 48.5 24.2 72.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.672 min Scan# 4356
Delta R.T. -0.005 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion:152 Resp: 718209
Ion Ratio Lower Upper
152 100
115 60.1 27.5 82.5
150 170.1 0.0 344.4









#73
Isopropylbenzene
Concen: 46.561 ug/l
RT: 12.578 min Scan# 3948
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 2975257
Ion Ratio Lower Upper
105 100
120 25.9 13.4 40.1



#74
N-ethyl acetate
Concen: 40.755 ug/l
RT: 12.387 min Scan# 3877
Delta R.T. -0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 43 Resp: 1151705
Ion Ratio Lower Upper
43 100
70 35.6 33.4 50.0
55 23.1 19.3 28.9
61 20.9 18.5 27.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 44.449 ug/l
 RT: 12.827 min Scan# 4041
 Delta R.T. 0.000 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55




Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion: 83 Resp: 982176
 Ion Ratio Lower Upper
 83 100
 131 10.3 5.5 16.5
 85 64.2 32.4 97.2



#76
 1,2,3-Trichloropropane
 Concen: 57.897 ug/l
 RT: 12.878 min Scan# 4060
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55


Tgt Ion: 75 Resp: 1192485
 Ion Ratio Lower Upper
 75 100
 77 131.4 82.6 247.8








#77
Bromobenzene
Concen: 45.850 ug/l
RT: 12.859 min Scan# 4053
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:156 Resp: 688172
Ion Ratio Lower Upper
156 100
77 227.6 90.6 271.8
158 97.2 48.9 146.7



#78
n-propylbenzene
Concen: 47.887 ug/l
RT: 12.918 min Scan# 4075
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 91 Resp: 3405115
Ion Ratio Lower Upper
91 100
120 21.8 11.9 35.9









#79
2-Chlorotoluene
Concen: 46.293 ug/l
RT: 13.007 min Scan# 4109
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 2079987
Ion Ratio Lower Upper
91 100
126 32.6 18.3 54.8



#80
1,3,5-Trimethylbenzene
Concen: 47.313 ug/l
RT: 13.058 min Scan# 4127
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55

Tgt Ion: 105 Resp: 2433427
Ion Ratio Lower Upper
105 100
120 47.5 24.6 73.6









#81
 trans-1,4-Dichloro-2-butene
 Concen: 39.726 ug/l
 RT: 12.623 min Scan# 3965
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 244287
 Ion Ratio Lower Upper
 75 100
 53 120.7 90.5 135.7
 89 46.3 38.2 57.4



#82
 4-Chlorotoluene
 Concen: 47.342 ug/l
 RT: 13.103 min Scan# 4144
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55


Tgt Ion: 91 Resp: 2000404
 Ion Ratio Lower Upper
 91 100
 126 32.3 18.2 54.6








#83
 tert-Butylbenzene
 Concen: 46.438 ug/l
 RT: 13.323 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55

Instrument :
 MSVOA_N
 ClientSampleId :



Tgt Ion:119 Resp: 2114866




Ion	Ratio	Lower	Upper
119	100		
91	70.1	31.5	94.5
134	26.1	12.4	37.4



#84
 1,2,4-Trimethylbenzene
 Concen: 48.727 ug/l
 RT: 13.366 min Scan# 4242
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55

Tgt Ion:105 Resp: 2386736

Ion	Ratio	Lower	Upper
105	100		
120	44.1	22.8	68.4









#85
 sec-Butylbenzene
 Concen: 47.151 ug/l
 RT: 13.500 min Scan# 41
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55




Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:105 Resp: 2912663
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 48.282 ug/l
 RT: 13.613 min Scan# 4334
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55

Tgt Ion:119 Resp: 2331317
 Ion Ratio Lower Upper
 119 100
 134 25.7 13.6 40.7
 91 25.1 11.4 34.1









#87
 1,3-Dichlorobenzene
 Concen: 47.043 ug/l
 RT: 13.616 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55




Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion:146 Resp: 1159987
 Ion Ratio Lower Upper
 146 100
 111 42.2 18.6 55.6
 148 64.0 31.6 95.0



#88
 1,4-Dichlorobenzene
 Concen: 46.451 ug/l
 RT: 13.694 min Scan# 4364
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55

Tgt Ion:146 Resp: 1125604
 Ion Ratio Lower Upper
 146 100
 111 40.9 18.3 54.8
 148 64.4 31.8 95.4









#89
n-Butylbenzene
Concen: 47.074 ug/l
RT: 13.940 min Scan# 4457
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 1843840
Ion Ratio Lower Upper
91 100
92 51.1 25.4 76.0
134 24.2 13.6 40.8



#90
Hexachloroethane
Concen: 47.734 ug/l
RT: 14.208 min Scan# 4556
Delta R.T. -0.003 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55


Tgt Ion: 117 Resp: 404664
Ion Ratio Lower Upper
117 100
201 76.8 43.9 131.6








#91
1,2-Dichlorobenzene
Concen: 47.191 ug/l
RT: 13.989 min Scan# 4474
Delta R.T. 0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:146 Resp: 1133784
Ion Ratio Lower Upper
146 100
111 43.0 19.1 57.3
148 64.0 32.1 96.3




#92
1,2-Dibromo-3-Chloropropane
Concen: 44.345 ug/l
RT: 14.603 min Scan# 4703
Delta R.T. 0.000 min
Lab File: VN069753.D
Acq: 4 Dec 2021 6:55




Tgt Ion: 75 Resp: 177064
Ion Ratio Lower Upper
75 100
155 74.7 46.9 140.7
157 94.9 60.6 181.7








#93
 1,2,4-Trichlorobenzene
 Concen: 43.049 ug/l
 RT: 15.262 min Scan# 4950
 Delta R.T. -0.003 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55


Instrument :
 MSVOA_N
 ClientSampleId :





Tgt Ion:180 Resp: 502264
 Ion Ratio Lower Upper
 180 100
 182 94.7 48.3 144.9
 145 31.2 14.5 43.5




#94
 Hexachlorobutadiene
 Concen: 43.883 ug/l
 RT: 15.370 min Scan# 4989
 Delta R.T. -0.000 min
 Lab File: VN069753.D
 Acq: 4 Dec 2021 6:55



Tgt Ion:225 Resp: 279413
 Ion Ratio Lower Upper
 225 100
 223 61.4 31.9 95.7
 227 62.8 32.1 96.3





#95

Naphthalene

Concen: 40.664 ug/l

RT: 15.504 min Scan# 5040

Delta R.T. -0.003 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:128 Resp: 1398885

Ion Ratio Lower Upper

128 100

127 13.5 10.5 15.7

129 10.9 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 43.216 ug/l

RT: 15.700 min Scan# 5112

Delta R.T. -0.002 min

Lab File: VN069753.D

Acq: 4 Dec 2021 6:55

Tgt Ion:180 Resp: 492299

Ion Ratio Lower Upper

180 100

182 95.8 47.4 142.2

145 32.6 15.0 45.1

