

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069719.D
 Acq On : 3 Dec 2021 15:38
 Operator : JC/MD
 Sample : M4922-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D4-20211129

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/04/2021
 Supervised By : Mahesh Dadoda 12/06/2021

Quant Time: Dec 04 05:32:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1467898	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2549530	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2611765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	997985	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1085911	47.91	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.82%		
35) Dibromofluoromethane	8.024	113	754132	47.22	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.44%		
50) Toluene-d8	10.443	98	3236269	50.47	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.94%		
62) 4-Bromofluorobenzene	12.733	95	1215703	52.33	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	104.66%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.075	85	5088	0.30	ug/l	93
9) 1,1,2-Trichlorotrifluo...	4.207	101	54362	3.75	ug/l	97
12) 1,1-Dichloroethene	4.175	96	10270	0.73	ug/l #	82
27) cis-1,2-Dichloroethene	7.329	96	24408	1.33	ug/l	98
44) Trichloroethene	9.217	130	2574683	145.41	ug/l	90
49) 1,4-Dioxane	9.574	88	5075	11.34	ug/l #	70
64) Tetrachloroethene	10.977	164	53343	3.01	ug/l	94
95) Naphthalene	15.509	128	4848	4.08	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069719.D
Acq On : 3 Dec 2021 15:38
Operator : JC/MD
Sample : M4922-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D4-20211129

Quant Time: Dec 04 05:32:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/04/2021
Supervised By :Mahesh Dadoda 12/06/2021

