

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069725.D
 Acq On : 3 Dec 2021 18:12
 Operator : JC/MD
 Sample : M4922-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE117D2-20211130

Quant Time: Dec 04 05:34:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1460348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2536163	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2573040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	958660	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1086798	48.200	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.400%
35) Dibromofluoromethane	8.024	113	752579	47.374	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.740%
50) Toluene-d8	10.443	98	3220860	50.494	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.980%
62) 4-Bromofluorobenzene	12.733	95	1179202	51.028	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.060%
Target Compounds						
16) Acetone	4.299	43	40794	3.008	ug/l	95
44) Trichloroethene	9.215	130	7288	0.414	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069725.D
Acq On : 3 Dec 2021 18:12
Operator : JC/MD
Sample : M4922-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE117D2-20211130

Quant Time: Dec 04 05:34:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

