

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069726.D
 Acq On : 3 Dec 2021 18:37
 Operator : JC/MD
 Sample : M4922-22
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D3-20211130

Quant Time: Dec 04 05:34:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1473482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2569138	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2592562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	959488	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1117844	49.135	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.260%		
35) Dibromofluoromethane	8.021	113	765059	47.541	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.080%		
50) Toluene-d8	10.443	98	3255752	50.386	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.780%		
62) 4-Bromofluorobenzene	12.733	95	1188068	50.752	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	101.500%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.205	101	25590	1.757	ug/l	98
16) Acetone	4.296	43	31865	2.328	ug/l	96
44) Trichloroethene	9.217	130	3616065	202.661	ug/l	92
64) Tetrachloroethene	10.982	164	5875	0.334	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069726.D
Acq On : 3 Dec 2021 18:37
Operator : JC/MD
Sample : M4922-22
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D3-20211130

Quant Time: Dec 04 05:34:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

