

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069727.D
 Acq On : 3 Dec 2021 19:03
 Operator : JC/MD
 Sample : M4922-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D5-20211129

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/04/2021
 Supervised By : Mahesh Dadoda 12/06/2021

Quant Time: Dec 04 05:35:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1391893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2422176	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2482941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	921602	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1023954	47.65	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.30%		
35) Dibromofluoromethane	8.024	113	717440	47.29	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.58%		
50) Toluene-d8	10.443	98	3107160	51.00	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.00%		
62) 4-Bromofluorobenzene	12.731	95	1134240	51.39	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	102.78%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.075	85	15143	0.93	ug/l	98
3) Chloromethane	2.306	50	5529m	0.27	ug/l	
9) 1,1,2-Trichlorotrifluo...	4.213	101	95971	6.98	ug/l	93
12) 1,1-Dichloroethene	4.186	96	19486	1.47	ug/l	95
24) 1,1-Dichloroethane	6.391	63	13484	0.45	ug/l	98
27) cis-1,2-Dichloroethene	7.329	96	18958	1.09	ug/l	97
30) Chloroform	7.823	83	22704	0.74	ug/l	99
44) Trichloroethene	9.218	130	1301183	77.35	ug/l	93
55) 1,1,2-Trichloroethane	10.896	97	5030	0.29	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069727.D
Acq On : 3 Dec 2021 19:03
Operator : JC/MD
Sample : M4922-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D5-20211129

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/04/2021
Supervised By : Mahesh Dadoda 12/06/2021

Quant Time: Dec 04 05:35:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

