

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069737.D
 Acq On : 4 Dec 2021 00:08
 Operator : JC/MD
 Sample : M4922-07 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D6-20211129

Quant Time: Dec 04 05:39:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/04/2021
 Supervised By : Mahesh Dadoda 12/06/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1338602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2324347	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2344065	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	832090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	993391	48.064	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.120%		
35) Dibromofluoromethane	8.024	113	684984	47.048	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.100%		
50) Toluene-d8	10.440	98	2965267	50.723	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.440%		
62) 4-Bromofluorobenzene	12.731	95	1046937	49.433	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.860%		
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.207	101	18267	1.381	ug/l #	73
12) 1,1-Dichloroethene	4.183	96	6681m	0.524	ug/l	
27) cis-1,2-Dichloroethene	7.329	96	3966	0.237	ug/l	82
44) Trichloroethene	9.217	130	1215165	75.276	ug/l	93
93) 1,2,4-Trichlorobenzene	15.265	180	9816	3.556	ug/l	95
95) Naphthalene	15.507	128	41939	4.934	ug/l	99
96) 1,2,3-Trichlorobenzene	15.697	180	11563	3.262	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069737.D
Acq On : 4 Dec 2021 00:08
Operator : JC/MD
Sample : M4922-07 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D6-20211129

Quant Time: Dec 04 05:39:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/04/2021
Supervised By :Mahesh Dadoda 12/06/2021

