

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069740.D
 Acq On : 4 Dec 2021 1:24
 Operator : JC/MD
 Sample : M4922-11 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20211130

Quant Time: Dec 04 05:40:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1375113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2399193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2511530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	939810	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1033438	48.674	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.340%
35) Dibromofluoromethane	8.024	113	716538	47.680	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.360%
50) Toluene-d8	10.440	98	3112387	51.579	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.160%
62) 4-Bromofluorobenzene	12.731	95	1166246	53.349	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.700%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	7787	0.610	ug/l #	78
44) Trichloroethene	9.217	130	216986	13.022	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069740.D
Acq On : 4 Dec 2021 1:24
Operator : JC/MD
Sample : M4922-11 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20211130

Quant Time: Dec 04 05:40:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

