

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069741.D
 Acq On : 4 Dec 2021 1:50
 Operator : JC/MD
 Sample : M4922-12 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP02-20211130

Quant Time: Dec 04 05:40:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1440906	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2552830	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2634685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1000481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1100529	49.467	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.940%
35) Dibromofluoromethane	8.021	113	756386	47.303	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.600%
50) Toluene-d8	10.440	98	3262749	50.816	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.640%
62) 4-Bromofluorobenzene	12.731	95	1227759	52.783	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.560%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.218	101	9512	0.668	ug/l #	49
44) Trichloroethene	9.215	130	169385	9.554	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069741.D
Acq On : 4 Dec 2021 1:50
Operator : JC/MD
Sample : M4922-12 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP02-20211130

Quant Time: Dec 04 05:40:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

