

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069742.D
 Acq On : 4 Dec 2021 2:15
 Operator : JC/MD
 Sample : M4922-14 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP06-20211130

Quant Time: Dec 04 05:40:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1491160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2610209	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2713436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1018868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1112928	48.339	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.680%
35) Dibromofluoromethane	8.021	113	773615	47.317	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.640%
50) Toluene-d8	10.441	98	3359003	51.165	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.340%
62) 4-Bromofluorobenzene	12.731	95	1246627	52.416	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.840%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.215	130	93764	5.172	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069742.D
Acq On : 4 Dec 2021 2:15
Operator : JC/MD
Sample : M4922-14 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP06-20211130

Quant Time: Dec 04 05:40:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

