

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069748.D
 Acq On : 4 Dec 2021 4:48
 Operator : JC/MD
 Sample : M4919-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-20211202

Quant Time: Dec 04 05:42:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1461451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2579395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2655927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1007695	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1129649	50.062	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.120%
35) Dibromofluoromethane	8.021	113	780708	48.321	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.640%
50) Toluene-d8	10.440	98	3298639	50.846	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.700%
62) 4-Bromofluorobenzene	12.731	95	1225002	52.122	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069748.D
Acq On : 4 Dec 2021 4:48
Operator : JC/MD
Sample : M4919-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-20211202

Quant Time: Dec 04 05:42:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

