

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069731.D
 Acq On : 3 Dec 2021 20:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 05:37:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.583	0.529	9.3	88	0.00
3 P	Chloromethane	0.731	0.633	13.4	91	0.00
4 C	Vinyl Chloride	0.672	0.598	11.0#	91	0.00
5 T	Bromomethane	0.365	0.324	11.2	90	0.00
6 T	Chloroethane	0.412	0.361	12.4	91	0.00
7 T	Trichlorofluoromethane	0.882	0.784	11.1	92	0.00
8 T	Diethyl Ether	0.340	0.300	11.8	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.442	10.5	92	0.00
10 T	Methyl Iodide	0.675	0.605	10.4	87	0.00
11 T	Tert butyl alcohol	0.146	0.106	27.4#	73	-0.01
12 CM	1,1-Dichloroethene	0.476	0.424	10.9#	92	0.00
13 T	Acrolein	0.124	0.097	21.8	85	0.00
14 T	Allyl chloride	1.073	0.934	13.0	89	0.00
15 T	Acrylonitrile	0.393	0.330	16.0	84	0.00
16 T	Acetone	0.464	0.307	33.8#	67	0.00
17 T	Carbon Disulfide	1.230	1.106	10.1	90	0.00
18 T	Methyl Acetate	1.402	1.132	19.3	84	0.00
19 T	Methyl tert-butyl Ether	1.852	1.634	11.8	90	0.00
20 T	Methylene Chloride	0.617	0.523	15.2	92	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.457	11.6	91	0.00
22 T	Diisopropyl ether	2.098	1.905	9.2	91	0.00
23 T	Vinyl Acetate	1.707	1.563	8.4	89	0.00
24 P	1,1-Dichloroethane	1.087	0.982	9.7	92	0.00
25 T	2-Butanone	0.631	0.486	23.0	77	0.00
26 T	2,2-Dichloropropane	0.845	0.699	17.3	85	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.558	10.7	91	0.00
28 T	Bromochloromethane	0.565	0.488	13.6	89	0.00
29 T	Tetrahydrofuran	0.383	0.319	16.7	82	0.00
30 C	Chloroform	1.097	0.990	9.8#	92	0.00
31 T	Cyclohexane	1.074	0.946	11.9	92	0.00
32 T	1,1,1-Trichloroethane	0.935	0.849	9.2	91	0.00
33 S	1,2-Dichloroethane-d4	0.772	0.665	13.9	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	110	0.00
35 S	Dibromofluoromethane	0.313	0.278	11.2	90	0.00
36 T	1,1-Dichloropropene	0.476	0.436	8.4	91	0.00
37 T	Ethyl Acetate	0.708	0.594	16.1	84	0.00
38 T	Carbon Tetrachloride	0.467	0.431	7.7	92	0.00
39 T	Methylcyclohexane	0.566	0.519	8.3	90	0.00
40 TM	Benzene	1.455	1.322	9.1	91	0.00
41 T	Methacrylonitrile	0.333	0.293	12.0	86	0.00
42 TM	1,2-Dichloroethane	0.587	0.524	10.7	91	0.00
43 T	Isopropyl Acetate	1.048	0.908	13.4	86	0.00
44 TM	Trichloroethene	0.347	0.321	7.5	92	0.00
45 C	1,2-Dichloropropane	0.391	0.351	10.2#	91	0.00
46 T	Dibromomethane	0.256	0.231	9.8	90	0.00
47 T	Bromodichloromethane	0.485	0.456	6.0	92	0.00
48 T	Methyl methacrylate	0.489	0.428	12.5	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069731.D
 Acq On : 3 Dec 2021 20:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 05:37:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.007	22.2	80	0.00
50 S	Toluene-d8	1.258	1.091	13.3	88	0.00
51 T	4-Methyl-2-Pentanone	0.680	0.599	11.9	84	0.00
52 CM	Toluene	0.888	0.819	7.8#	91	0.00
53 T	t-1,3-Dichloropropene	0.510	0.475	6.9	87	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.519	5.3	88	0.00
55 T	1,1,2-Trichloroethane	0.358	0.321	10.3	89	0.00
56 T	Ethyl methacrylate	0.608	0.556	8.6	88	0.00
57 T	1,3-Dichloropropane	0.631	0.565	10.5	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.154	0.138	10.4	85	0.00
59 T	2-Hexanone	0.514	0.427	16.9	79	0.00
60 T	Dibromochloromethane	0.340	0.324	4.7	90	0.00
61 T	1,2-Dibromoethane	0.371	0.337	9.2	89	0.00
62 S	4-Bromofluorobenzene	0.456	0.394	13.6	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.339	0.310	8.6	91	0.00
65 PM	Chlorobenzene	1.014	0.923	9.0	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.336	5.9	90	0.00
67 C	Ethyl Benzene	1.843	1.734	5.9#	90	0.00
68 T	m/p-Xylenes	0.671	0.634	5.5	90	0.00
69 T	o-Xylene	0.676	0.633	6.4	90	0.00
70 T	Styrene	1.072	1.040	3.0	90	0.00
71 P	Bromoform	0.264	0.252	4.5	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
73 T	Isopropylbenzene	4.449	3.876	12.9	91	0.00
74 T	N-amyl acetate	1.967	1.625	17.4	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.538	1.265	17.8	87	0.00
76 T	1,2,3-Trichloropropane	1.434	1.153	19.6	83	0.00
77 T	Bromobenzene	1.045	0.883	15.5	89	0.00
78 T	n-propylbenzene	4.950	4.398	11.2	90	0.00
79 T	2-Chlorotoluene	3.128	2.710	13.4	91	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.169	11.5	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.396	0.338	14.6	81	0.00
82 T	4-Chlorotoluene	2.942	2.578	12.4	90	0.00
83 T	tert-Butylbenzene	3.171	2.756	13.1	89	0.00
84 T	1,2,4-Trimethylbenzene	3.410	3.089	9.4	90	0.00
85 T	sec-Butylbenzene	4.301	3.856	10.3	90	0.00
86 T	p-Isopropyltoluene	3.362	3.072	8.6	91	0.00
87 T	1,3-Dichlorobenzene	1.717	1.498	12.8	90	0.00
88 T	1,4-Dichlorobenzene	1.687	1.450	14.0	90	0.00
89 T	n-Butylbenzene	2.727	2.455	10.0	88	0.00
90 T	Hexachloroethane	0.590	0.535	9.3	89	0.00
91 T	1,2-Dichlorobenzene	1.673	1.451	13.3	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.226	18.7	82	0.00
93 T	1,2,4-Trichlorobenzene	0.760	0.625	17.8	83	0.00
94 T	Hexachlorobutadiene	0.443	0.376	15.1	85	0.00
95 T	Naphthalene	2.199	1.637	25.6#	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069731.D
Acq On : 3 Dec 2021 20:45
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 04 05:37:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.612	18.3	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6