

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069756.D
 Acq On : 4 Dec 2021 8:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 09:52:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.583	0.627	-7.5	93	0.00
3 P	Chloromethane	0.731	0.690	5.6	88	0.00
4 C	Vinyl Chloride	0.672	0.686	-2.1#	93	0.00
5 T	Bromomethane	0.365	0.258	29.3#	64	0.01
6 T	Chloroethane	0.412	0.399	3.2	89	0.01
7 T	Trichlorofluoromethane	0.882	0.889	-0.8	93	0.01
8 T	Diethyl Ether	0.340	0.352	-3.5	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.506	-2.4	94	0.00
10 T	Methyl Iodide	0.675	0.465	31.1#	60	0.00
11 T	Tert butyl alcohol	0.146	0.122	16.4	75	-0.02
12 CM	1,1-Dichloroethene	0.476	0.498	-4.6#	96	0.00
13 T	Acrolein	0.124	0.080	35.5#	63	0.00
14 T	Allyl chloride	1.073	1.080	-0.7	92	0.00
15 T	Acrylonitrile	0.393	0.367	6.6	84	0.00
16 T	Acetone	0.464	0.337	27.4#	65	0.00
17 T	Carbon Disulfide	1.230	1.373	-11.6	100	0.00
18 T	Methyl Acetate	1.402	1.288	8.1	86	0.00
19 T	Methyl tert-butyl Ether	1.852	1.914	-3.3	94	0.00
20 T	Methylene Chloride	0.617	0.604	2.1	95	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.529	-2.3	94	0.00
22 T	Diisopropyl ether	2.098	2.227	-6.1	95	0.00
23 T	Vinyl Acetate	1.707	1.835	-7.5	93	0.00
24 P	1,1-Dichloroethane	1.087	1.136	-4.5	95	0.00
25 T	2-Butanone	0.631	0.537	14.9	76	0.00
26 T	2,2-Dichloropropane	0.845	0.722	14.6	78	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.636	-1.8	92	0.00
28 T	Bromochloromethane	0.565	0.557	1.4	91	0.00
29 T	Tetrahydrofuran	0.383	0.352	8.1	81	0.00
30 C	Chloroform	1.097	1.119	-2.0#	93	0.00
31 T	Cyclohexane	1.074	1.079	-0.5	94	0.00
32 T	1,1,1-Trichloroethane	0.935	0.984	-5.2	94	0.00
33 S	1,2-Dichloroethane-d4	0.772	0.778	-0.8	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.313	0.320	-2.2	94	0.00
36 T	1,1-Dichloropropene	0.476	0.489	-2.7	93	0.00
37 T	Ethyl Acetate	0.708	0.648	8.5	83	0.00
38 T	Carbon Tetrachloride	0.467	0.482	-3.2	93	0.00
39 T	Methylcyclohexane	0.566	0.584	-3.2	92	0.00
40 TM	Benzene	1.455	1.473	-1.2	93	0.00
41 T	Methacrylonitrile	0.333	0.330	0.9	87	0.00
42 TM	1,2-Dichloroethane	0.587	0.585	0.3	92	0.00
43 T	Isopropyl Acetate	1.048	1.038	1.0	89	0.00
44 TM	Trichloroethene	0.347	0.362	-4.3	95	0.00
45 C	1,2-Dichloropropane	0.391	0.402	-2.8#	95	0.00
46 T	Dibromomethane	0.256	0.260	-1.6	92	0.00
47 T	Bromodichloromethane	0.485	0.511	-5.4	93	0.00
48 T	Methyl methacrylate	0.489	0.482	1.4	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
 Data File : VN069756.D
 Acq On : 4 Dec 2021 8:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 04 09:52:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.007	22.2	71	0.00
50 S	Toluene-d8	1.258	1.258	0.0	92	0.00
51 T	4-Methyl-2-Pentanone	0.680	0.658	3.2	83	0.00
52 CM	Toluene	0.888	0.914	-2.9#	92	0.00
53 T	t-1,3-Dichloropropene	0.510	0.535	-4.9	90	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.579	-5.7	90	0.00
55 T	1,1,2-Trichloroethane	0.358	0.358	0.0	90	0.00
56 T	Ethyl methacrylate	0.608	0.618	-1.6	89	0.00
57 T	1,3-Dichloropropane	0.631	0.639	-1.3	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.154	0.161	-4.5	90	0.00
59 T	2-Hexanone	0.514	0.472	8.2	79	0.00
60 T	Dibromochloromethane	0.340	0.369	-8.5	93	0.00
61 T	1,2-Dibromoethane	0.371	0.373	-0.5	89	0.00
62 S	4-Bromofluorobenzene	0.456	0.457	-0.2	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.339	0.354	-4.4	94	0.00
65 PM	Chlorobenzene	1.014	1.038	-2.4	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.377	-5.6	91	0.00
67 C	Ethyl Benzene	1.843	1.950	-5.8#	91	0.00
68 T	m/p-Xylenes	0.671	0.711	-6.0	91	0.00
69 T	o-Xylene	0.676	0.712	-5.3	91	0.00
70 T	Styrene	1.072	1.165	-8.7	91	0.00
71 P	Bromoform	0.264	0.282	-6.8	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.449	4.383	1.5	92	0.00
74 T	N-amyl acetate	1.967	1.985	-0.9	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.538	1.436	6.6	88	0.00
76 T	1,2,3-Trichloropropane	1.434	1.267	11.6	82	0.00
77 T	Bromobenzene	1.045	1.004	3.9	91	0.00
78 T	n-propylbenzene	4.950	4.994	-0.9	91	0.00
79 T	2-Chlorotoluene	3.128	3.055	2.3	92	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.570	0.3	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.396	0.263	33.6#	57	0.00
82 T	4-Chlorotoluene	2.942	2.942	0.0	92	0.00
83 T	tert-Butylbenzene	3.171	3.108	2.0	90	0.00
84 T	1,2,4-Trimethylbenzene	3.410	3.511	-3.0	92	0.00
85 T	sec-Butylbenzene	4.301	4.324	-0.5	90	0.00
86 T	p-Isopropyltoluene	3.362	3.454	-2.7	91	0.00
87 T	1,3-Dichlorobenzene	1.717	1.704	0.8	91	0.00
88 T	1,4-Dichlorobenzene	1.687	1.652	2.1	92	0.00
89 T	n-Butylbenzene	2.727	2.813	-3.2	90	0.00
90 T	Hexachloroethane	0.590	0.525	11.0	78	0.00
91 T	1,2-Dichlorobenzene	1.673	1.663	0.6	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.264	5.0	85	0.00
93 T	1,2,4-Trichlorobenzene	0.760	0.791	-4.1	94	0.00
94 T	Hexachlorobutadiene	0.443	0.417	5.9	85	0.00
95 T	Naphthalene	2.199	2.337	-6.3	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120321\
Data File : VN069756.D
Acq On : 4 Dec 2021 8:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 04 09:52:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.785	-4.8	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6