

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120325\
 Data File : VN088421.D
 Acq On : 03 Dec 2025 12:31
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 04 00:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	171012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	392624	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	373269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	172595	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	186044	57.750	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.500%
35) Dibromofluoromethane	8.147	113	138712	50.981	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.960%
50) Toluene-d8	10.541	98	474572	46.877	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.760%
62) 4-Bromofluorobenzene	12.829	95	169931	48.920	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120325\
Data File : VN088421.D
Acq On : 03 Dec 2025 12:31
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 04 00:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

