

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120325\
 Data File : VN088425.D
 Acq On : 03 Dec 2025 13:53
 Operator : JC\MD
 Sample : Q3756-03 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BR-01-190-120225

Quant Time: Dec 04 00:18:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	174817	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	411266	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	398817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	182896	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	198155	60.171	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 120.340%		
35) Dibromofluoromethane	8.147	113	146765	51.495	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.000%		
50) Toluene-d8	10.541	98	501799	47.320	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 94.640%		
62) 4-Bromofluorobenzene	12.829	95	177307	48.730	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 97.460%		

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120325\
Data File : VN088425.D
Acq On : 03 Dec 2025 13:53
Operator : JC\MD
Sample : Q3756-03 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BR-01-190-120225

Quant Time: Dec 04 00:18:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

