

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120325\
 Data File : VN088430.D
 Acq On : 03 Dec 2025 15:35
 Operator : JC\MD
 Sample : Q3750-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FENCE WC-2

Quant Time: Dec 04 00:20:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	191162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	447334	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	477758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	223026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	233732	64.906	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 129.820%#		
35) Dibromofluoromethane	8.147	113	162894	52.546	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.100%		
50) Toluene-d8	10.541	98	584955	50.714	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.420%		
62) 4-Bromofluorobenzene	12.829	95	214794	54.273	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 108.540%		
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	34000	26.107	ug/l	97
18) Methyl Acetate	5.030	43	5324	1.373	ug/l #	73
20) Methylene Chloride	5.271	84	10785	3.871	ug/l #	91
25) 2-Butanone	7.477	43	37610	17.964	ug/l	90
43) Isopropyl Acetate	8.677	43	20537	2.137	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120325\
Data File : VN088430.D
Acq On : 03 Dec 2025 15:35
Operator : JC\MD
Sample : Q3750-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FENCE WC-2

Quant Time: Dec 04 00:20:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

