

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
 Data File : VN052648.D
 Acq On : 4 Dec 2018 12:27
 Operator : MD\SY
 Sample : J6209-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-12-3-18

Quant Time: Dec 05 10:15:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1150527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1779270	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1571337	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	599215	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	608542	48.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.16%	
35) Dibromofluoromethane	7.59	113	553550	50.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.92%	
50) Toluene-d8	10.09	98	2157573	50.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.06%	
62) 4-Bromofluorobenzene	12.40	95	690508	48.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
Data File : VN052648.D
Acq On : 4 Dec 2018 12:27
Operator : MD\SY
Sample : J6209-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TB-12-3-18

Quant Time: Dec 05 10:15:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

