

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
 Data File : VN052654.D
 Acq On : 4 Dec 2018 15:07
 Operator : MD\SY
 Sample : PB115293TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115293TB

Quant Time: Dec 05 03:56:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1203512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1857692	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1623941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	573636	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	641195	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
35) Dibromofluoromethane	7.59	113	575489	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
50) Toluene-d8	10.09	98	2266768	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	12.40	95	686597	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds						
16) Acetone	3.82	43	20527	5.939	ug/l	95
20) Methylene Chloride	4.55	84	32950	2.436	ug/l	96
43) Isopropyl Acetate	8.17	43	52919	2.921	ug/l #	82
95) Naphthalene	15.14	128	12276	3.958	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
Data File : VN052654.D
Acq On : 4 Dec 2018 15:07
Operator : MD\SY
Sample : PB115293TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115293TB

Quant Time: Dec 05 03:56:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

