

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
 Data File : VN052655.D
 Acq On : 4 Dec 2018 15:34
 Operator : MD\SY
 Sample : PB115293ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115293ZHE#01

Quant Time: Dec 05 03:58:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1199529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1857876	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1648725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	637407	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	637055	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
35) Dibromofluoromethane	7.59	113	570655	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
50) Toluene-d8	10.09	98	2266477	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
62) 4-Bromofluorobenzene	12.40	95	722426	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
Target Compounds						
16) Acetone	3.82	43	17800	5.167	ug/l	91
20) Methylene Chloride	4.55	84	86179	6.393	ug/l	98
43) Isopropyl Acetate	8.17	43	52740	2.911	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
Data File : VN052655.D
Acq On : 4 Dec 2018 15:34
Operator : MD\SY
Sample : PB115293ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115293ZHE#01

Quant Time: Dec 05 03:58:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

