

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
 Data File : VN052660.D
 Acq On : 4 Dec 2018 17:47
 Operator : MD\SY
 Sample : J6192-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 HD-02-120318

Quant Time: Dec 05 04:12:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1176841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1819219	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1602609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	608792	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	625864	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
35) Dibromofluoromethane	7.59	113	562386	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	10.09	98	2224311	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.40	95	697428	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
16) Acetone	3.82	43	32433	9.596	ug/l	96
20) Methylene Chloride	4.55	84	1767216	133.619	ug/l	100
43) Isopropyl Acetate	8.17	43	57694	3.252	ug/l #	76
95) Naphthalene	15.13	128	39279	5.025	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
Data File : VN052660.D
Acq On : 4 Dec 2018 17:47
Operator : MD\SY
Sample : J6192-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
HD-02-120318

Quant Time: Dec 05 04:12:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

