

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
 Data File : VN052662.D
 Acq On : 4 Dec 2018 18:41
 Operator : MD\SY
 Sample : J6201-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11932-72-11967

Quant Time: Dec 05 04:24:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1206530	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1870053	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1603449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	550982	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	633752	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
35) Dibromofluoromethane	7.59	113	569971	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	10.09	98	2264742	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	12.40	95	672663	44.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.22%	
Target Compounds						
16) Acetone	3.82	43	19586	5.653	ug/l	95
20) Methylene Chloride	4.55	84	214848	15.845	ug/l	99
43) Isopropyl Acetate	8.17	43	62038	3.401	ug/l #	76
95) Naphthalene	15.14	128	8853	3.826	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN120418\
Data File : VN052662.D
Acq On : 4 Dec 2018 18:41
Operator : MD\SY
Sample : J6201-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
72-11932-72-11967

Quant Time: Dec 05 04:24:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

