

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120419\
 Data File : VN059485.D
 Acq On : 4 Dec 2019 14:30
 Operator : JC/SP
 Sample : K6084-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30740

Quant Time: Dec 05 07:14:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	348107	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	530839	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	470110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	178523	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	206239	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
35) Dibromofluoromethane	7.57	113	160653	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
50) Toluene-d8	10.08	98	646763	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	12.40	95	207218	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.76	59	13273	16.288	ug/l #	73
16) Acetone	3.78	43	255547	102.755	ug/l #	57
20) Methylene Chloride	4.52	84	11891	1.762	ug/l #	62
25) 2-Butanone	6.81	43	30501	9.280	ug/l #	89
31) Cyclohexane	7.63	56	780199	107.088	ug/l	99
39) Methylcyclohexane	9.07	83	241391	36.312	ug/l	99
40) Benzene	8.02	78	391706	24.215	ug/l	100
43) Isopropyl Acetate	8.15	43	74366	8.103	ug/l #	86
52) Toluene	10.15	92	81264	8.306	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120419\
Data File : VN059485.D
Acq On : 4 Dec 2019 14:30
Operator : JC/SP
Sample : K6084-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30740

Quant Time: Dec 05 07:14:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

