


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN120419\
Data File : VN059475.D
Acq On : 4 Dec 2019 9:21
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Title : SW846 8260
Last Update : Tue Dec 03 16:37:42 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	10956	PASS
75	95	30	60	49.1	29405	PASS
95	95	100	100	100.0	59912	PASS
96	95	5	9	7.6	4544	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.6	46506	PASS
175	174	5	9	6.5	3044	PASS
176	174	95	101	98.5	45805	PASS
177	176	5	9	6.6	3040	PASS