

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN120420\
 Data File : VN064910.D
 Acq On : 4 Dec 2020 11:23
 Operator : JC/MD
 Sample : VN1204MBL01
 Misc : 5.00G/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1204MBL01

Quant Time: Dec 05 01:52:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	171923	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	301982	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	315224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	139937	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.98	65	104436	52.25	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.50%	
35) Dibromofluoromethane		7.55	113	96699	53.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.02%	
50) Toluene-d8		10.06	98	380209	51.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.00%	
62) 4-Bromofluorobenzene		12.37	95	146818	55.57	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120420\
Data File : VN064910.D
Acq On : 4 Dec 2020 11:23
Operator : JC/MD
Sample : VN1204MBL01
Misc : 5.00G/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1204MBL01

Quant Time: Dec 05 01:52:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

