

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120420\
 Data File : VN064915.D
 Acq On : 4 Dec 2020 13:40
 Operator : JC/MD
 Sample : L4936-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-214

Quant Time: Dec 05 02:11:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	175906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	307447	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	330586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	148016	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	106768	52.21	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	104.42%	
35) Dibromofluoromethane	7.55	113	96962	52.71	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	105.42%	
50) Toluene-d8	10.06	98	394312	52.46	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	104.92%	
62) 4-Bromofluorobenzene	12.38	95	156722	58.26	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	116.52%	

Target Compounds

					Qvalue
16) Acetone	3.78	43	20451	30.421 ug/l	99
20) Methylene Chloride	4.50	84	10888	5.169 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120420\
Data File : VN064915.D
Acq On : 4 Dec 2020 13:40
Operator : JC/MD
Sample : L4936-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-214

Quant Time: Dec 05 02:11:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

