

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120420\
 Data File : VN064918.D
 Acq On : 4 Dec 2020 14:53
 Operator : JC/MD
 Sample : L4954-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS013-05-568-120120

Quant Time: Dec 05 02:20:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	164927	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	293506	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	305921	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	134268	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	103204	53.83	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	107.66%	
35) Dibromofluoromethane	7.55	113	96347	54.86	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	109.72%	
50) Toluene-d8	10.06	98	367694	51.24	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	102.48%	
62) 4-Bromofluorobenzene	12.38	95	139854	54.46	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	108.92%	

Target Compounds					Qvalue
16) Acetone	3.78	43	2971	4.714 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120420\
Data File : VN064918.D
Acq On : 4 Dec 2020 14:53
Operator : JC/MD
Sample : L4954-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS013-05-568-120120

Quant Time: Dec 05 02:20:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

