

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120420\
 Data File : VN064926.D
 Acq On : 4 Dec 2020 18:06
 Operator : JC/MD
 Sample : L4954-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS013-13-FD01-120120

Quant Time: Dec 05 03:06:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	166689	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	298401	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	305696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	134957	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102332	52.81	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	105.62%	
35) Dibromofluoromethane	7.55	113	96327	53.95	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	107.90%	
50) Toluene-d8	10.06	98	367395	50.36	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	100.72%	
62) 4-Bromofluorobenzene	12.38	95	141891	54.34	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	108.68%	

Target Compounds					Qvalue
16) Acetone	3.78	43	1754	2.753 ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120420\
Data File : VN064926.D
Acq On : 4 Dec 2020 18:06
Operator : JC/MD
Sample : L4954-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS013-13-FD01-120120

Quant Time: Dec 05 03:06:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

