

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
 Data File : VN069759.D
 Acq On : 4 Dec 2021 12:11
 Operator : JC/MD
 Sample : VN1204MBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1204MBL01

Quant Time: Dec 04 16:10:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1420388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2487568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2535599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	965312	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1066382	48.625	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.240%
35) Dibromofluoromethane	8.021	113	741302	47.575	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.160%
50) Toluene-d8	10.440	98	3153565	50.404	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.800%
62) 4-Bromofluorobenzene	12.731	95	1191980	52.589	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
Data File : VN069759.D
Acq On : 4 Dec 2021 12:11
Operator : JC/MD
Sample : VN1204MBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1204MBL01

Quant Time: Dec 04 16:10:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

