

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN120421\  
 Data File : VN069764.D  
 Acq On : 4 Dec 2021 14:40  
 Operator : JC/MD  
 Sample : M4922-10DL 5X  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 RE131D3-20211130DL

Quant Time: Dec 04 16:11:53 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N120221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Dec 03 08:46:47 2021  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|------------------------------|--------|-------|----------|----------|--------|----------|
| -----                        |        |       |          |          |        |          |
| Internal Standards           |        |       |          |          |        |          |
| 1) Pentafluorobenzene        | 8.088  | 168   | 1356839  | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.968  | 114   | 2390754  | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.744 | 117   | 2460283  | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.678 | 152   | 929432   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds  |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.440  | 65    | 1026534  | 49.000   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 61 - 141 | Recovery | =      | 98.000%  |
| 35) Dibromofluoromethane     | 8.024  | 113   | 710288   | 47.431   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 69 - 133 | Recovery | =      | 94.860%  |
| 50) Toluene-d8               | 10.443 | 98    | 3056882  | 50.838   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 65 - 126 | Recovery | =      | 101.680% |
| 62) 4-Bromofluorobenzene     | 12.734 | 95    | 1134006  | 52.057   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 58 - 135 | Recovery | =      | 104.120% |
| Target Compounds             |        |       |          |          |        |          |
|                              |        |       |          |          | Qvalue |          |
| 9) 1,1,2-Trichlorotrifluo... | 4.213  | 101   | 585844   | 43.680   | ug/l   | 96       |
| 44) Trichloroethene          | 9.217  | 130   | 62920    | 3.789    | ug/l   | 94       |
| 64) Tetrachloroethene        | 10.977 | 164   | 21027    | 1.260    | ug/l   | 93       |
| -----                        |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN120421\  
Data File : VN069764.D  
Acq On : 4 Dec 2021 14:40  
Operator : JC/MD  
Sample : M4922-10DL 5X  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
RE131D3-20211130DL

Quant Time: Dec 04 16:11:53 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N120221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Dec 03 08:46:47 2021  
Response via : Initial Calibration

