

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
 Data File : VN069765.D
 Acq On : 4 Dec 2021 15:05
 Operator : JC/MD
 Sample : M4922-15DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE131D1-20211130DL

Quant Time: Dec 04 16:12:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1374568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2424774	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2506577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	959066	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1028243	48.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.90%
35) Dibromofluoromethane	8.024	113	715307	47.10	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.20%
50) Toluene-d8	10.443	98	3112259	51.03	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.06%
62) 4-Bromofluorobenzene	12.733	95	1171352	53.02	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.04%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.210	101	12786	0.94	ug/l #	71
27) cis-1,2-Dichloroethene	7.324	96	10613	0.62	ug/l	97
44) Trichloroethene	9.217	130	545652	32.40	ug/l	94
64) Tetrachloroethene	10.982	164	32286	1.90	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
Data File : VN069765.D
Acq On : 4 Dec 2021 15:05
Operator : JC/MD
Sample : M4922-15DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE131D1-20211130DL

Quant Time: Dec 04 16:12:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

