

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
 Data File : VN069766.D
 Acq On : 4 Dec 2021 15:31
 Operator : JC/MD
 Sample : M4922-22DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D3-20211130DL

Quant Time: Dec 04 16:12:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1492667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2605941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2637775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	1008746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1128502	48.966	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.940%
35) Dibromofluoromethane	8.024	113	780961	47.844	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.680%
50) Toluene-d8	10.443	98	3303013	50.395	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.800%
62) 4-Bromofluorobenzene	12.731	95	1228588	51.742	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.480%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.208	101	5747	0.390	ug/l #	79
44) Trichloroethene	9.218	130	693378	38.311	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
Data File : VN069766.D
Acq On : 4 Dec 2021 15:31
Operator : JC/MD
Sample : M4922-22DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D3-20211130DL

Quant Time: Dec 04 16:12:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

