

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
 Data File : VN069769.D
 Acq On : 4 Dec 2021 16:47
 Operator : JC/MD
 Sample : M4922-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP06-20211130

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/06/2021
 Supervised By : Mahesh Dadoda 12/06/2021

Quant Time: Dec 06 00:17:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1343922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2375619	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2429434	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	900673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1027588	49.522	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.040%		
35) Dibromofluoromethane	8.021	113	709460	47.678	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.360%		
50) Toluene-d8	10.443	98	3016853	50.492	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.980%		
62) 4-Bromofluorobenzene	12.731	95	1114922	51.507	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	103.020%		
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.208	101	38505	2.898	ug/l	96
12) 1,1-Dichloroethene	4.181	96	26495	2.070	ug/l	97
16) Acetone	4.299	43	21283	1.705	ug/l	92
24) 1,1-Dichloroethane	6.385	63	10763m	0.369	ug/l	
27) cis-1,2-Dichloroethene	7.321	96	16666	0.992	ug/l	84
30) Chloroform	7.815	83	14181	0.481	ug/l #	81
38) Carbon Tetrachloride	8.206	117	16736	0.755	ug/l	87
44) Trichloroethene	9.218	130	1626207	98.565	ug/l	91
55) 1,1,2-Trichloroethane	10.899	97	10457	0.614	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
Data File : VN069769.D
Acq On : 4 Dec 2021 16:47
Operator : JC/MD
Sample : M4922-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP06-20211130

Quant Time: Dec 06 00:17:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/06/2021
Supervised By :Mahesh Dadoda 12/06/2021

