

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
 Data File : VN069773.D
 Acq On : 4 Dec 2021 18:30
 Operator : JC/MD
 Sample : M4922-13DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP04-20211130DL

Quant Time: Dec 06 00:18:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1309079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2331009	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2447020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	939374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1011241	50.031	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.060%	
35) Dibromofluoromethane	8.024	113	700724	47.992	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 95.980%	
50) Toluene-d8	10.443	98	3002627	51.215	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.440%	
62) 4-Bromofluorobenzene	12.731	95	1153357	54.302	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.600%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.213	101	4644	0.359	ug/l #	55
27) cis-1,2-Dichloroethene	7.321	96	9730	0.595	ug/l	88
44) Trichloroethene	9.217	130	516943	31.932	ug/l	92
64) Tetrachloroethene	10.977	164	29740	1.792	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120421\
Data File : VN069773.D
Acq On : 4 Dec 2021 18:30
Operator : JC/MD
Sample : M4922-13DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP04-20211130DL

Quant Time: Dec 06 00:18:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

