

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080331.D
 Acq On : 04 Dec 2023 12:23
 Operator : JC\MD
 Sample : PB157524TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157524TB

Quant Time: Dec 05 04:13:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153359	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	278658	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	320868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133352	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	138260	54.601	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.200%
35) Dibromofluoromethane	8.171	113	108622	52.828	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.660%
50) Toluene-d8	10.571	98	389255	52.517	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.040%
62) 4-Bromofluorobenzene	12.853	95	159160	56.851	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.700%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	16187	19.032	ug/l	90
20) Methylene Chloride	5.289	84	10985	6.168	ug/l #	91
25) 2-Butanone	7.483	43	16303	16.134	ug/l #	83
37) Ethyl Acetate	7.577	43	3029	1.467	ug/l #	71
43) Isopropyl Acetate	8.688	43	87244	21.712	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080331.D
Acq On : 04 Dec 2023 12:23
Operator : JC\MD
Sample : PB157524TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157524TB

Quant Time: Dec 05 04:13:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

