

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080334.D
 Acq On : 04 Dec 2023 13:35
 Operator : JC\MD
 Sample : PB157524ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157524ZHE#03

Quant Time: Dec 05 04:14:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	151881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	293928	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	319276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	135842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	139546	55.645	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.280%	
35) Dibromofluoromethane	8.171	113	110534	50.965	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.940%	
50) Toluene-d8	10.571	98	396590	50.727	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.460%	
62) 4-Bromofluorobenzene	12.853	95	153565	52.003	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	104.000%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	20526	24.369	ug/l #	83
20) Methylene Chloride	5.277	84	35603	20.184	ug/l	93
43) Isopropyl Acetate	8.694	43	124924	29.474	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080334.D
Acq On : 04 Dec 2023 13:35
Operator : JC\MD
Sample : PB157524ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157524ZHE#03

Quant Time: Dec 05 04:14:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

