

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080335.D
 Acq On : 04 Dec 2023 13:58
 Operator : JC\MD
 Sample : PB157524ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157524ZHE#04

Quant Time: Dec 05 04:15:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	149399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	291136	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	312865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	131946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	137882	55.895	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.780%			
35) Dibromofluoromethane	8.165	113	108916	50.701	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.400%			
50) Toluene-d8	10.571	98	392594	50.698	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.400%			
62) 4-Bromofluorobenzene	12.853	95	148396	50.734	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.460%			
Target Compounds						
16) Acetone	4.430	43	21112	25.481	ug/l	98
20) Methylene Chloride	5.277	84	32318	18.626	ug/l	94
43) Isopropyl Acetate	8.688	43	133423	31.782	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080335.D
Acq On : 04 Dec 2023 13:58
Operator : JC\MD
Sample : PB157524ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157524ZHE#04

Quant Time: Dec 05 04:15:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

