

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080336.D
 Acq On : 04 Dec 2023 14:22
 Operator : JC\MD
 Sample : PB157524ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157524ZHE#05

Quant Time: Dec 05 04:15:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	303281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	315893	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	131009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	139949	55.177	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.360%
35) Dibromofluoromethane	8.171	113	110809	49.517	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.040%
50) Toluene-d8	10.570	98	397955	49.332	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.853	95	152789	50.144	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.280%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	17399	20.424	ug/l #	83
20) Methylene Chloride	5.283	84	11540	6.469	ug/l	95
25) 2-Butanone	7.494	43	16210	16.015	ug/l	92
43) Isopropyl Acetate	8.694	43	146721	33.550	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080336.D
Acq On : 04 Dec 2023 14:22
Operator : JC\MD
Sample : PB157524ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157524ZHE#05

Quant Time: Dec 05 04:15:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

