

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080337.D
 Acq On : 04 Dec 2023 14:46
 Operator : JC\MD
 Sample : PB157524ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157524ZHE#06

Quant Time: Dec 05 04:16:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	155830	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293828	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	318382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	141775	55.101	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.200%
35) Dibromofluoromethane	8.171	113	110864	51.135	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.280%
50) Toluene-d8	10.570	98	396874	50.781	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.560%
62) 4-Bromofluorobenzene	12.853	95	154667	52.394	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.780%
Target Compounds						
						Qvalue
16) Acetone	4.441	43	16813	19.455	ug/l	98
20) Methylene Chloride	5.277	84	11905	6.578	ug/l	92
25) 2-Butanone	7.482	43	15904	15.489	ug/l	94
43) Isopropyl Acetate	8.688	43	140806	33.233	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080337.D
Acq On : 04 Dec 2023 14:46
Operator : JC\MD
Sample : PB157524ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157524ZHE#06

Quant Time: Dec 05 04:16:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

