

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080341.D
 Acq On : 04 Dec 2023 16:21
 Operator : JC\MD
 Sample : 05602-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPRING-VALLEY-TP-2

Quant Time: Dec 05 04:17:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	155071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	292383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	323040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	142596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	139737	54.575	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.140%		
35) Dibromofluoromethane	8.171	113	110995	51.448	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.900%		
50) Toluene-d8	10.571	98	400027	51.437	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.880%		
62) 4-Bromofluorobenzene	12.853	95	158346	53.905	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 107.800%		
Target Compounds						
16) Acetone	4.430	43	20704	24.074	ug/l	Qvalue 91
20) Methylene Chloride	5.277	84	17472	9.701	ug/l	91
43) Isopropyl Acetate	8.688	43	119375	28.314	ug/l #	81
95) Naphthalene	15.647	128	190129	23.405	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080341.D
Acq On : 04 Dec 2023 16:21
Operator : JC\MD
Sample : 05602-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPRING-VALLEY-TP-2

Quant Time: Dec 05 04:17:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

