

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080342.D
 Acq On : 04 Dec 2023 16:45
 Operator : JC\MD
 Sample : 05602-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPRING-VALLEY-TP-3

Quant Time: Dec 05 04:18:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	152567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	295345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	327288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	142425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	141114	56.017	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 112.040%		
35) Dibromofluoromethane	8.171	113	111468	51.150	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.300%		
50) Toluene-d8	10.571	98	404481	51.488	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.980%		
62) 4-Bromofluorobenzene	12.853	95	163601	55.135	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 110.280%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	20099	23.755	ug/l #	77
20) Methylene Chloride	5.277	84	17932	10.120	ug/l #	89
43) Isopropyl Acetate	8.688	43	135143	31.733	ug/l #	78
95) Naphthalene	15.647	128	41567	5.123	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080342.D
Acq On : 04 Dec 2023 16:45
Operator : JC\MD
Sample : 05602-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPRING-VALLEY-TP-3

Quant Time: Dec 05 04:18:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

