

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080343.D
 Acq On : 04 Dec 2023 17:09
 Operator : JC\MD
 Sample : 05602-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ROUTE-4-TP-4

Quant Time: Dec 05 04:18:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	150319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	290851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	318925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	139541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	139311	56.128	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.260%
35) Dibromofluoromethane	8.165	113	110605	51.538	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.080%
50) Toluene-d8	10.571	98	395802	51.162	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.320%
62) 4-Bromofluorobenzene	12.853	95	158575	54.267	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.540%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	27048	32.445	ug/l	94
20) Methylene Chloride	5.283	84	17371	9.950	ug/l	97
43) Isopropyl Acetate	8.688	43	124028	29.573	ug/l #	79
95) Naphthalene	15.641	128	18107	2.278	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080343.D
Acq On : 04 Dec 2023 17:09
Operator : JC\MD
Sample : 05602-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ROUTE-4-TP-4

Quant Time: Dec 05 04:18:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

