

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080344.D
 Acq On : 04 Dec 2023 17:33
 Operator : JC\MD
 Sample : 05602-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ROUTE-4-TP-5

Quant Time: Dec 05 04:19:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	157355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	307049	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	329347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	139478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	143678	55.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.600%
35) Dibromofluoromethane	8.165	113	112848	49.809	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.620%
50) Toluene-d8	10.571	98	409314	50.117	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.240%
62) 4-Bromofluorobenzene	12.853	95	159025	51.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.100%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	41316	47.345	ug/l #	88
20) Methylene Chloride	5.277	84	17053	9.331	ug/l #	92
25) 2-Butanone	7.489	43	8583	8.278	ug/l	89
43) Isopropyl Acetate	8.694	43	124901	28.210	ug/l #	77
95) Naphthalene	15.641	128	11597	1.460	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080344.D
Acq On : 04 Dec 2023 17:33
Operator : JC\MD
Sample : 05602-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ROUTE-4-TP-5

Quant Time: Dec 05 04:19:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

