

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080345.D
 Acq On : 04 Dec 2023 17:57
 Operator : JC\MD
 Sample : 05602-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ROUTE-4-TP-6

Quant Time: Dec 05 04:20:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	155043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	296099	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	326507	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	143765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	143192	55.934	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.860%
35) Dibromofluoromethane	8.171	113	112235	51.370	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.740%
50) Toluene-d8	10.571	98	406385	51.599	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.200%
62) 4-Bromofluorobenzene	12.853	95	159008	53.451	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.900%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	39639	46.100	ug/l	98
20) Methylene Chloride	5.277	84	17377	9.650	ug/l #	87
25) 2-Butanone	7.488	43	8197	8.024	ug/l #	74
43) Isopropyl Acetate	8.694	43	131650	30.834	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080345.D
Acq On : 04 Dec 2023 17:57
Operator : JC\MD
Sample : 05602-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ROUTE-4-TP-6

Quant Time: Dec 05 04:20:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

