

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080346.D
 Acq On : 04 Dec 2023 18:20
 Operator : JC\MD
 Sample : 05602-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPRING-VALLEY-TP-1

Quant Time: Dec 05 04:20:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	153749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	298518	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	326687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	142573	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	145083	57.150	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.300%			
35) Dibromofluoromethane	8.171	113	114021	51.765	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.520%			
50) Toluene-d8	10.571	98	404244	50.911	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.820%			
62) 4-Bromofluorobenzene	12.853	95	155908	51.984	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 103.960%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	35630	41.786	ug/l	94
20) Methylene Chloride	5.277	84	16630	9.313	ug/l	93
43) Isopropyl Acetate	8.694	43	99779	23.180	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080346.D
Acq On : 04 Dec 2023 18:20
Operator : JC\MD
Sample : 05602-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPRING-VALLEY-TP-1

Quant Time: Dec 05 04:20:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

