

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
 Data File : VN080349.D
 Acq On : 04 Dec 2023 19:32
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 05 04:22:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	170011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	311420	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	324741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	147584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	146347	52.134	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.260%			
35) Dibromofluoromethane	8.171	113	117506	51.137	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.280%			
50) Toluene-d8	10.571	98	400657	48.369	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.740%			
62) 4-Bromofluorobenzene	12.853	95	159350	50.931	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.860%			
Target Compounds						
					Qvalue	
40) Benzene	8.606	78	26537	2.807	ug/l	95
95) Naphthalene	15.647	128	212797	25.310	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120423\
Data File : VN080349.D
Acq On : 04 Dec 2023 19:32
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 05 04:22:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

