



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN120423\  
Data File : VN080325.D  
Acq On : 04 Dec 2023 09:08  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N111623W.M  
Title : SW846 8260  
Last Update : Fri Nov 17 01:19:32 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1843

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 17.2         | 7974       | PASS                |
| 75             | 95              | 30              | 60              | 55.4         | 25624      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 46264      | PASS                |
| 96             | 95              | 5               | 9               | 6.9          | 3205       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.2          | 474        | PASS                |
| 174            | 95              | 50              | 100             | 83.1         | 38456      | PASS                |
| 175            | 174             | 5               | 9               | 7.3          | 2825       | PASS                |
| 176            | 174             | 95              | 101             | 95.4         | 36693      | PASS                |
| 177            | 176             | 5               | 9               | 7.1          | 2596       | PASS                |

Average of 12.847 to 12.859 min.: VN080325.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 35.90 | 334    | 50.00 | 7974   | 65.10 | 322    | 78.95 | 211    |
| 37.10 | 2048   | 51.05 | 2686   | 68.00 | 4972   | 79.80 | 18     |
| 38.10 | 1841   | 52.00 | 52     | 69.05 | 5338   | 80.00 | 44     |
| 39.00 | 859    | 56.05 | 640    | 69.95 | 433    | 80.90 | 224    |
| 44.05 | 232    | 56.95 | 1126   | 71.90 | 182    | 81.90 | 458    |
| 45.05 | 439    | 59.90 | 537    | 73.05 | 2482   | 86.95 | 1612   |
| 46.20 | 87     | 60.95 | 2280   | 74.00 | 8900   | 87.95 | 1811   |
| 46.85 | 368    | 61.95 | 2373   | 75.05 | 25624  | 90.80 | 87     |
| 47.85 | 274    | 63.00 | 2321   | 76.00 | 2059   | 91.00 | 114    |
| 48.20 | 122    | 64.00 | 453    | 76.85 | 369    | 91.95 | 1755   |
| 49.00 | 2029   | 64.90 | 231    | 78.20 | 121    | 93.00 | 2515   |

Average of 12.847 to 12.859 min.: VN080325.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 94.05  | 5831   | 117.85 | 302    | 141.90 | 79     | 173.05 | 474    |
| 95.00  | 46264  | 118.85 | 484    | 142.80 | 1024   | 173.95 | 38456  |
| 96.05  | 3205   | 127.90 | 184    | 143.10 | 217    | 174.90 | 2825   |
| 96.80  | 66     | 128.95 | 224    | 144.85 | 454    | 175.90 | 36693  |
| 103.90 | 356    | 129.80 | 183    | 145.70 | 55     | 176.95 | 2596   |
| 105.00 | 57     | 132.80 | 87     | 148.00 | 60     | 191.05 | 189    |
| 105.75 | 269    | 133.10 | 170    | 154.85 | 228    | 192.95 | 404    |
| 106.00 | 60     | 134.60 | 104    | 157.05 | 102    | 206.80 | 116    |
| 108.10 | 52     | 135.00 | 99     | 171.80 | 94     | 207.10 | 72     |
| 115.80 | 151    | 139.70 | 78     | 172.05 | 288    | 248.70 | 86     |
| 116.95 | 752    | 140.90 | 882    | 172.90 | 211    | 249.20 | 66     |

Average of 12.847 to 12.859 min.: VN080325.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 264.90 | 128    |     |        |     |        |     |        |
| 266.80 | 115    |     |        |     |        |     |        |
| 267.05 | 168    |     |        |     |        |     |        |
| 269.20 | 55     |     |        |     |        |     |        |
| 281.10 | 2240   |     |        |     |        |     |        |
| 282.10 | 499    |     |        |     |        |     |        |
| 282.90 | 358    |     |        |     |        |     |        |
| 283.10 | 75     |     |        |     |        |     |        |
| 325.15 | 120    |     |        |     |        |     |        |
| 340.80 | 201    |     |        |     |        |     |        |
| 341.10 | 50     |     |        |     |        |     |        |

Instrument :

MSVOA\_N

ClientSampleId :

BFB